

UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING	DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/797,391	03/10/2004		Eleanor Schuler	920607-95597	4322
7590 01/03/2006			EXAM	EXAMINER	
Francis Law Group 1942 Embarcadero				ALTER, ALYSSA M	
Oakland, CA 94606			ART UNIT	PAPER NUMBER	
				3762	
				DATE MAIL ED: 01/03/2004	6

Please find below and/or attached an Office communication concerning this application or proceeding.

<u></u>	Application No.	Applicant(s)				
	10/797,391	SCHULER ET AL.				
Office Action Summary	Examiner	Art Unit				
	Alyssa M. Alter	3762				
The MAILING DATE of this communication app Period for Reply		orrespondence address				
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely. - If NO period for reply signified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).						
Status						
1)⊠ Responsive to communication(s) filed on <u>30 June 2005</u> .						
,—	,— ,,					
closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.						
Disposition of Claims						
4) Claim(s) <u>1-17</u> is/are pending in the application.						
4a) Of the above claim(s) is/are withdrawn from consideration.						
5) Claim(s) is/are allowed.						
6)⊠ Claim(s) <u>1-17</u> is/are rejected.						
7) Claim(s) is/are objected to.	a alaatiaa aassiissaa t					
8) Claim(s) are subject to restriction and/or election requirement.						
Application Papers						
9)⊠ The specification is objected to by the Examine	r.					
10)⊠ The drawing(s) filed on <u>10 March 2004</u> is/are: a)⊠ accepted or b)□ objected to by the Examiner.						
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).						
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).						
11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.						
Priority under 35 U.S.C. § 119						
12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of:						
1. Certified copies of the priority documents have been received.						
2. Certified copies of the priority documents have been received in Application No						
3. Copies of the certified copies of the priority documents have been received in this National Stage						
application from the International Bureau (PCT Rule 17.2(a)).						
* See the attached detailed Office action for a list of the certified copies not received.						
	•					
	-					
Attachment(s)	4) 🔲 Interview Summary	(PTO-413)				
1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948)	Paper No(s)/Mail D	ate				
3) Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) Paper No(s)/Mail Date	5) Notice of Informal F	Patent Application (PTO-152)				
S Patent and Trademark Office						

Application/Control Number: 10/797,391

Art Unit: 3762

DETAILED ACTION

Terminal Disclaimer

The terminal disclaimers filed on June 30, 2005 disclaiming the terminal portion of any patent granted on this application, which would extend beyond the expiration date of US Application 10/781,078 and US Patent 6,681,136 have been reviewed and is accepted. The terminal disclaimers have been recorded. Therefore the Double Patenting rejections in view of co-pending US Application 10/781, 078 and US Patent 6,681,136 have been withdrawn.

Response to Arguments

Applicant's arguments with respect to claims 1-5, 13 and 15-16 have been – considered but are moot in view of the new ground(s) of rejection. Applicant's arguments regarding claims 6-12, 14 and 17 have been fully considered but they are not persuasive. Claims 1-5, 13 and 15-16 are rejected under new subject matter and claims 6-12, 14 and 17 are rejected over Kieval et al. (US 6,522,926).

Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

1. Claims 1-5, 13 and 15-16 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter, which was not described in the specification in such a way as to reasonably

convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. While noted that the other application 10/000, 005 is incorporated by reference, it is noted that incorporation by reference of essential subject matter is improper (see 37 CFR 1.57). While the specification discloses a plurality of waveforms, the original disclosure does not provide support for a second waveform signal from/following the first waveform signal and transmitting it to the body, in combination with the other steps in the claim.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

- (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.
- Claims 6 and 10-12 stands rejected and claims 14 and 17 are rejected under 35 U.S.C. 102(e) as anticipated by or, in the alternative, under 35 U.S.C. 103(a) as obvious over Kieval et al. (US 6,522,926) for reasons disclosed in the previous Office Action. Kieval et al. discloses a device to "be used to increase or decrease blood pressure." sympathetic nervous system activity and neurohormonal activity, as needed to minimize deleterious effects on the heart, vasculature and other organs and tissues" (col. 21, lines 11-14) by activating the baroreceptors. Kieval et al. also discloses in column 21, lines 15-16 that "the baroreceptor activation devices described previously may also be used to provide antiarrhythmic effects". As seen in figure 3, "the control system 60 generates

Application/Control Number: 10/797,391

Art Unit: 3762

a control signal as a function of the received sensor signal. The control signal activates, deactivates or otherwise modulates the baroreceptor activation device 70. Typically, activation of the device 70 results in activation of the baroreceptors 30"(col. 9, lines 33-37). The examiner considers the control system to be the storage area where the signals are generated.

In the alternative, Kieval et al. discloses the claimed invention except for the memory to store waveforms. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the control system and method as taught by Kieval et al. with a memory to store waveforms since it was known in the art that storing and recording data can provide physicians with information on the status of their patient.

As to claim 6, the Applicant merely states a source of collected waveform signals indicative of a body organ functioning. Since Kieval et al. generates a control signal as a function of the received sensor signal, then the signals are indicative of body organ functioning. Furthermore, Kieval et al. does "directly" transmit waveform signals as disclosed in col. 7, lines 33-37.

- 2. Claims 7 and 8 stand rejected under 35 U.S.C. 103(a) as being unpatentable over Kieval et al. (US 6,522,926) for reasons previously made of record in the previous Office Action.
- 3. Claim 9 stands rejected under 35 U.S.C. 102(e) as anticipated by or, in the alternative, under 35 U.S.C. 103(a) as obvious over Kieval et al. (US 6,522,926) for reasons previously made of record in the previous Office Action.

Specification

The specification is objected to as failing to provide proper antecedent basis for the claimed subject matter. See 37 CFR 1.75(d)(1) and MPEP § 608.01(o). Correction of the following is required: a second waveform signal following/from the first waveform signal.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Art Unit: 3762

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Alyssa M. Alter whose telephone number is (571) 272-4939. The examiner can normally be reached on M-F 9am to 4pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on (571) 272-4955. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

> Wyma M. attor Alyssa M Alter Examiner

Art Unit 3762